

GN300 Datsheet



AQL 1.5

EN455 Parts 1, 2, 3 & 4

MDD 93/42/EEC

EN374

Class 1 Medical Device



General Details

Material	Acrylonitrile Butadiene (Nitrile), non sterile
Surface Finish	Powder Free
Colour	Blue
Geometry	Ambidextrous
Texture	Finger Textured
Cuff and Finishing	Beaded

Glove Specifications - EN455

Size	Length	Palm Width	Cuff	Thickness	
	mm	mm		Palm	Finger
XS	240	75 ± 10	Min. 0.05	Min. 0.05	Min. 0.05
S	240	80 ± 10	Min. 0.05	Min. 0.05	Min. 0.05
M	240	95 ± 10	Min. 0.05	Min. 0.05	Min. 0.05
L	240	110 ± 10	Min. 0.05	Min. 0.05	Min. 0.05
XL	240	≥ 110	Min. 0.05	Min. 0.05	Min. 0.05

Performance Requirements

Test	Related Defects	Inspection Level	AQL
Water Leak at Final Release	Holes	G-1 Single Sampling Plan	1.5

Manufacturing Standards

Maunufacturing Standards	Incompliance with ISO 9001 and ISO 13485
Presentation	Standard: 200 gloves per dispenser box; 10 dispenser boxes per Carton
Glove Care	Should not store under direct sunlight or at high heat and humid conditions. Single use only.

Stretch Nitrile Blue Powder Free Gloves



G·S·L

WWW.GLOVEMAN.CO.UK

Gloveman Supplies Ltd

Vision House, Treleigh Ind Est,

Redruth TR16 4AX

t. 01209 314759

f. 01209 314760